

	<h1>TARAline BR1</h1>
indicator	bromine
Application	Drinking water, swimming pool water, service water, process water, sea water
bromine agents	Free bromine (HOBr) 1-Bromo-3-chloro-5.5-dimethyl-hydantoin (BCDMH)
Measuring system	membrane covered, amperometric potentiostatic 3-electrode system
electronic	<p>Analog version:</p> <ul style="list-style-type: none"> <li>- voltage output</li> <li>- not galvanically isolated electronics</li> <li>- analog internal data processing</li> <li>- output signal: analog (analog-out/analog)</li> </ul> <p>Digital version:</p> <ul style="list-style-type: none"> <li>- electronic is completely galvanically isolated</li> <li>- digital internal data processing</li> <li>- output signal: analog (analog-out/digital) or digital (digital-out/digital)</li> </ul> <p>mA-version:</p> <ul style="list-style-type: none"> <li>- current output analog</li> <li>- not galvanically isolated electronics</li> <li>- output signal: analog (analog-out/analog)</li> </ul>
Information about the measuring range of sensors with 4-20 mA	<p>Slope of a sensor can vary production-related or application-related between 65% and 150% of the nominal slope</p> <p>-&gt; Recommendation to determine the suitable measuring range or the suitable sensor: Concentration to be measured x factor 1.5 = measuring range of the sensor</p> <p>Example: Concentration to be measured 1.6 ppm x 1.5 = 2.4 -&gt; recommended sensor with a measuring range of 5 ppm</p>
Working temperature	<p>Measuring water temperature: 0 ... +45 °C (no ice crystals in the measuring water)</p> <p>Ambient temperature: 0 ... +55 °C</p>
Temperature compensation	Automatically, by an integrated temperature sensor Sudden temperature changes must be avoided
max. allowed working pressure	<p>Operation without retaining ring: 0.5 bar, no pressure impulses and/or vibrations</p> <p>Operation with retaining ring: 0.5 bar, no pressure impulses and/or vibrations</p>
Flow rate	approx. 15-30 L/h in TARAflow FLC
pH-range	pH 6.5 – pH 9.5, highly reduced dependence on pH – value (see diagram last page “relative dependence on pH”)
Run-in time	First start-up approx. 2 h
Response time	T <sub>90</sub> : approx. 2 min


	<h1>TARAline BR1</h1>
Zero point adjustment	Not necessary
Slope calibration	At the device, by analytical determination of the bromine concentration Recommendation depending on bromine agent: - Free bromine                   DPD1 - method - BCDMH                            DPD4 - method
Cross sensitivities/ interferences	Cl <sub>2</sub> :    is also measured ClO <sub>2</sub> :  is also measured O <sub>3</sub> :     is also measured  Corrosion inhibitors can lead to measuring errors. Stabilisers for water hardness can lead to measuring errors.
Absence of the disinfectant	Max. 24 h
Connection	analog-out/analog version:    4-pole plug adapter analog-out/digital version:   4-pole plug adapter digital-out/digital version:   5-pole M12, plug-on flange 4-20 mA version:                2-pole terminal or 5-pole M12, plug-on flange
material	Microporous hydrophilic membrane, PVC, PEEK ,stainless steel 1.4571
Size	diameter:                            approx. 25 mm Length: BR1 (analog-out/analog)    approx. 175 mm BR1 (analog-out/digital)    approx. 195 mm BR1 (digital-out/digital)    approx. 205 mm BR1 4-20 mA                    approx. 220 mm (2-pole-terminal) approx. 190 mm (5-pole-M12)
Transport	+5 ... +50 °C (Sensor, electrolyte, membrane cap)
storage	Sensor:                    dry and without electrolyte no limit at +5 ... +40 °C
	Electrolyte:            in original bottle protected from sunlight at +5 ... +35 °C min. 1 year or until the specified EXP-Date
	Membrane cap:        in original packing no limit at +5 ... +40 °C (used membrane caps can not be stored)
maintenance	Regularly control of the measuring signal, min. once a week The following information highly depends on the water quality: Change of the membrane cap:       once a year Change of the electrolyte:            every 3 - 6 months
	EMC-Testing DIN EN 61326-1, 61326-2-3 RoHS compliant

## Technical Data

### 1. BR1 (analog output, analog internal signal processing)

analog-out / analog

A potential-free electrical connection is necessary as the sensor electronic is not equipped with a galvanical isolation.


	Measuring range in ppm	Resolution in ppm	Output Output resistance	Nominal slope (at pH 7.2) in mV/ppm	Voltage supply	Connection
BR1H	0.005...2.000	0.001	analog 0...-2000 mV	-1000	±5 - ±15 VDC	4-pole screw connector
BR1N	0.05...20.00	0.01	1 kΩ	-100	10 mA	

(Subject to technical changes!)

### 2. BR1 (analog output, digital internal signal processing)

analog-out / digital

- The power supply is galvanically isolated inside of the sensor.
- The output signal is galvanically isolated too, that means potential-free.


	Measuring range in ppm	Resolution in ppm	Output Output resistance	Nominal Slope (at pH 7.2) in mV/ppm	Power supply	Connection
BR1H-An	0.005...2.000	0.001	analog 0...-2 V (max. -2.5 V)	-1000	9-30 VDC approx. 56-20 mA	4-pole screw connector
BR1N-An	0.05...20.00	0.01	1 kΩ	-100		
BR1H-Ap	0.005...2.000	0.001	analog 0...+2 V (max. +2.5 V)	+1000		
BR1N-Ap	0.05...20.00	0.01	1 kΩ	+100		

(Subject to technical changes!)

3. BR1 (digital output, digital internal signal processing)

digital-out / digital

- The power supply is galvanically isolated inside of the sensor.
- The output signal is galvanically isolated too, that means potential-free.

	Measuring range	Resolution	Output Output resistance	Power supply	Connection
	in ppm	in ppm			
BR1H-M0c	0.005...2.000	0.001	Modbus RTU	9-30 VDC	5-pole M12 plug-on flange
BR1N-M0c	0.05...20.00	0.01	There are no terminating resistors in the sensor.	approx. 56-20 mA	


(Subject to technical changes!)

4. BR1 4-20 mA (analog output, analog internal signal processing)

analog-out / analog


A potential-free electrical connection is necessary as the sensor electronic is not equipped with a galvanical isolation.

4.1 Electrical connection: 2 pole terminal clamp

	Measuring range	Resolution	Output Output resistance	Nominal slope (at pH 7.2)	Voltage supply	Connection
	in ppm	in ppm		in mA/ppm		
BR1MA-2	0.005 ... 2.000	0.001	analog 4...20 mA uncalibrated	8.0	12...30 VDC  R <sub>L</sub> = 50Ω (12V)... 900Ω (30V)	2-pole terminal (2 x 1 mm <sup>2</sup> )  Recommended: Round cable ∅ 4 mm 2 x 0.34 mm <sup>2</sup>
BR1MA-5	0.05 ... 5.00	0.01		3.2		
BR1MA-10	0.05 ... 10.00	0.01		1.6		
BR1MA-20	0.05 ... 20.00	0.01		0.8		

(Subject to technical changes!)

4.2 Electrical connection: 5 pole M12 plug-on flange

	Measuring range in ppm	Resolution in ppm	Output Output resistance	Nominal slope (at pH 7.2) in mA/ppm	Voltage supply	Connection
BR1MA-2-M12	0.005 ... 2.000	0.001	analog 4...20 mA uncalibrated	8.0	12...30 VDC  R <sub>L</sub> = 50Ω (12V)... 900Ω (30V)	5-pole M12 plug-on flange  Function of wires: PIN2: +U PIN3: -U
BR1MA-5-M12	0.05 ... 5.00	0.01		3.2		
BR1MA-10-M12	0.05 ... 10.00	0.01		1.6		
BR1MA-20-M12	0.05 ... 20.00	0.01		0.8		

(Subject to technical changes!)

**Spare Parts**

Type	Membrane cap	Electrolyte	Emery	O-ring
All BR1	M48.2 Art. No. 11047	ECP1.4/GEL, 100 ml Art. No. 11006.1	S1 Art. No. 11908	14 x 1.8 NBR Art. No. 11806

(Subject to technical changes!)

**relative dependence on pH**

Temperature: 25°C / Flow rate: 30 l/h

