


	<h1>TARAbase CL4.2</h1>						
indicator	Free chlorine, pH-dependent						
Application	Swimming pool water, drinking water, service water, process water The water must not contain any surfactants (tensides)! pH-value must be constant.						
Chlorination agents	inorganic chlorine compounds: NaOCl (=sodium hypochlorite), Ca(OCl) ₂ , chlorine gas, chlorine electrolysis with membrane cell (unsuitable: chlorine electrolysis without membrane cell)						
Measuring system	Membrane covered, amperometric 2-electrode system with electronic inside						
Electronic	<p>Analog version:</p> <ul style="list-style-type: none"> - voltage output - not galvanically isolated electronics - analog internal data processing <p>Digital version:</p> <ul style="list-style-type: none"> - output signal: analog (analog-out/analog) - electronic is completely galvanically isolated - digital internal data processing - output signal: analog (analog-out/digital) or digital (digital-out/digital) <p>mA-version:</p> <ul style="list-style-type: none"> - current output analog - not galvanically isolated electronics - output signal: analog (analog-out/analog) 						
Information about the measuring range	<p>The actual slope of a sensor can vary production-related between 65% and 150% of the nominal slope</p> <p>Note: With a slope > 100% the measuring range is reduced accordingly. (Ex.: 150% slope → 67% of the specified measuring range)</p>						
Accuracy after calibration at repeatability conditions (25°C, pH 7.2 in drinking water) of the upper full scale	<table border="0" style="width: 100%;"> <tr> <td style="width: 40%;">– Measuring range 2 mg/l:</td> <td style="width: 30%;">at 0.4 mg/l</td> <td style="width: 30%; text-align: right;"><1%</td> </tr> <tr> <td></td> <td>at 1.6 mg/l</td> <td style="text-align: right;"><1%</td> </tr> </table>	– Measuring range 2 mg/l:	at 0.4 mg/l	<1%		at 1.6 mg/l	<1%
– Measuring range 2 mg/l:	at 0.4 mg/l	<1%					
	at 1.6 mg/l	<1%					
Slope drift At repeatability conditions (25 °C, pH 7,2 in drinking water)	approx. <-1% per month						
Working temperature	Measuring water temperature: 0 ... +45 °C (no ice crystals in the measuring water)						
	Ambient temperature: 0 ... +55 °C						
Temperature compensation	Automatically, by an integrated temperature sensor Sudden temperature changes must be avoided						


	<h1>TARAbase CL4.2</h1>	
Max. allowed working pressure	Operation without retaining ring: 0.5 bar, no pressure impulses and/or vibrations	
	Operation with retaining ring: 1.0 bar, no pressure impulses and/or vibrations	
Flow rate (Incoming flow velocity)	approx. 15-30/h (33 – 66 cm/s) in TARAflow FLC, small flow rate dependence is given (see diagram "Slope of TARAbase CL4 versus flowrate", p. 8)	
pH-range	pH 6 – pH 8, pay attention to the dissociation equilibrium HOCL (see diagram "Slope of TARAbase CL4 versus pH, p. 8)	
Run-in time	First start-up approx. 1 h	
Response time	T ₉₀ : approx. 30 sec.	
Zero point adjustment	Not necessary	
Slope calibration	At the device, by analytical determination DPD-1-Method	
Interferences	ClO ₂ : factor 9 O ₃ Electrolytically generated chlorine with a cell without membrane can produce trouble	
Absence of the disinfectant	Max. 24 h	
Connection	analog-out/analog version: 4-pole plug adapter analog-out/digital version: 4-pole plug adapter digital-out/digital version: 5-pole M12, plug-on flange 4-20 mA version: 2-pole terminal or 5-pole M12, plug-on flange	
max. length of sensor cable (depending on internal signal processing)	analog	< 30 m
	digital	> 30 m are permissible Maximum cable length depends on application
material	Semipermeable membrane, PVC-U, ABS	
Size	diameter: approx. 25 mm Length: analog-out/analog version approx. 175 mm analog-out/digital version approx. 195 mm digital-out/digital version approx. 205 mm 4-20 mA version approx. 220 mm (2-pole-terminal) approx. 190 mm (5-pole-M12)	
Transport	+5 ... +50 °C (sensor, electrolyte, membrane cap)	

	<h1>TARAbase CL4.2</h1>	
<p>storage</p>	<p>Sensor:</p>	<p>dry and without electrolyte no limit at +5 ... +40 °C</p>
	<p>Electrolyte:</p>	<p>in original bottle protected from sunlight at +5 ... +35 °C min. 1 year or until specified EXP-Date</p>
	<p>Membrane cap:</p>	<p>in original packing no limit at +5 ... +40 °C (used membrane caps can not be stored)</p>
<p>maintenance</p>	<p>Regularly control of the measuring signal, min. once a week The following specifications depend on the water quality: Change of the membrane cap: once a year Change of the electrolyte: every 3 - 6 months</p>	
	<p>EMC-Testing DIN EN 61326-1, 61326-2-3, 63000 RoHS compliant</p>	

Technical Data
1. CL4.2 (analog output, analog internal signal processing)

analog-out / analog

A potential-free electrical connection is necessary as the sensor electronic is not equipped with a galvanical isolation.


	Measuring range in ppm	resolution in ppm	Output Output resistance	Nominal slope (at pH 7.2) in mV/ppm	Voltage supply	Connection
CL4.2N	0.05...20.00	0.01	0...-2000 mV 1 kΩ	-100	±5 - ±15 VDC 10 mA	4-pole screw connector
CL4.2H	0.005...2.000	0.001		-1000		
CL4.2DW	0.005...5.000	0.001		-300		
CL4.2L	0.5...200.0	0.1		-10		
CL4.2HUp	0.005...2.000	0.01	0...+2000 mV 1 kΩ	+1000	10 - 30 VDC 10 mA	
CL4.2Up	0.05...20.00	0.01		+100		

(Subject to technical changes!)

2. CL4.2 (analog output, digital internal signal processing)

analog-out / digital

- The power supply is galvanically isolated inside of the sensor.
- The output signal is galvanically isolated too, that means potential-free.


	Measuring range in ppm	resolution in ppm	Output Output resistance	Nominal slope (at pH 7.2) in mV/ppm	Power supply	Connection
CL4.2H-An	0.005...2.000	0.001	analog 0...-2 V (max. -2.5 V) 1 kΩ	-1000	9-30 VDC approx. 56-20 mA	4-pole screw connector
CL4.2N-An	0.05...20.00	0.01		-100		
CL4.2L-An	0.5...200.0	0.1		-10		
CL4.2H-Ap	0.005...2.000	0.001	analog 0...+2 V (max. +2.5 V) 1 kΩ	+1000		
CL4.2N-Ap	0.05...20.00	0.01		+100		
CL4.2L-Ap	0.5...200.0	0.1		+10		

(Subject to technical changes!)

3. CL4.2 (digital output, digital internal signal processing)

digital-out / digital

- The power supply is galvanically isolated inside of the sensor.
- The output signal is galvanically isolated too, that means potential-free.

	Measuring range in ppm	resolution in ppm	Output Output resistance	Power supply	Connection
CL4.2H-M0c	0.005... 2.000	0.001	Modbus RTU There are no terminating resistors in the sensor.	9-30 VDC approx. 56-20 mA	5-pole M12 plug-on flange
CL4.2N-M0c	0.05... 20.00	0.01			
CL4.2L-M0c	0.5...200.0	0.1			


(Subject to technical changes!)

4. CL4.2 4-20 mA (analog output, analog internal signal processing)

analog-out / analog


A potential-free electrical connection is necessary as the sensor electronic is not equipped with a galvanical isolation.

4.1 Electrical connection: 2 pole terminal clamp

	Measuring range	Resolution	Output Output resistance	Nominal slope (at pH 7.2)	Voltage supply	Connection
	in ppm	in ppm		in mA/ppm		
CL4.2MA0.5	0.005...0.500	0.001	4...20 mA uncalibrated	32.0	12...30 VDC R _L 50Ω...R _L 900Ω	2-pole terminal (2 x 1 mm ²) Recommended: Round cable ∅ 4 mm 2 x 0.34 mm ²
CL4.2MA2	0.005...2.000	0.001		8.0		
CL4.2MA5	0.05...5.00	0.01		3.2		
CL4.2MA10	0.05...10.00	0.01		1.6		
CL4.2MA20	0.05...20.00	0.01		0.8		
CL4.2MA-100	0.5...100.0	0.1		0.16		
CL4.2MA-200	0.5...200.0	0.1		0.8		

(Subject to technical changes!)

4.2 Electrical connection: 5 pole M12 plug-on flange

	Measuring range	resolution	Output Output resistance	Nominal slope (at pH 7.2)	Voltage supply	Connection
	in ppm	in ppm		in mA/ppm		
CL4.2MA0.5-M12	0.005...0.500	0.001	4...20 mA uncalibrated	32.0	12...30 VDC R _L 50Ω...R _L 900Ω	5-pole M12 plug-on flange Function of wires: PIN2: +U PIN3: -U
CL4.2MA2-M12	0.005...2.000	0.001		8.0		
CL4.2MA5-M12	0.05...5.00	0.01		3.2		
CL4.2MA10-M12	0.05...10.00	0.01		1.6		
CL4.2MA20-M12	0.05...20.00	0.01		0.8		
CL4.2MA-100-M12	0.5...100.0	0.1		0.16		
CL4.2MA-200-M12	0.5...200.0	0.1		0.8		

(Subject to technical changes!)

Spare Parts

Type	Membrane cap	Electrolyte	Emery	O-ring
For all CL4.2	M20.2 Art. no. 11011.1	ECL1, 100 ml Art. no. 11001	S1 Art. no. 11908	14 x 1.8 NBR Art. No. 11806

(Subject to technical changes!)

