

	<h1>TARAline BR1</h1>
indicator	bromine
Application	Drinking water, swimming pool water, service water, process water, sea water
bromine agents	Free bromine (HOBr) 1-Bromo-3-chloro-5.5-dimethyl-hydantoin (BCDMH)
Measuring system	membrane covered, amperometric potentiostatic 3-electrode system
electronic	<p>Analog version:</p> <ul style="list-style-type: none"> - voltage output - not galvanically isolated electronics - analog internal data processing - output signal: analog (analog-out/analog) <p>Digital version:</p> <ul style="list-style-type: none"> - electronic is completely galvanically isolated - digital internal data processing - output signal: analog (analog-out/digital) or digital (digital-out/digital) <p>mA-version:</p> <ul style="list-style-type: none"> - current output analog - not galvanically isolated electronics - output signal: analog (analog-out/analog)
Information about the measuring range	<p>The actual slope of a sensor can vary production-related between 65% and 150% of the nominal slope</p> <p>Note: With a slope > 100% the measuring range is reduced accordingly. (Ex.: 150% slope → 67% of the specified measuring range)</p>
Working temperature	Measuring water temperature: 0 ... +45 °C (no ice crystals in the measuring water)
	Ambient temperature: 0 ... +55 °C
Temperature compensation	Automatically, by an integrated temperature sensor Sudden temperature changes must be avoided
max. allowed working pressure	Operation without retaining ring: 0.5 bar, no pressure impulses and/or vibrations
	Operation with retaining ring: 0.5 bar, no pressure impulses and/or vibrations
Flow rate (Incoming flow velocity)	approx. 15-30 L/h (33 – 66 cm/s) in TARAflow FLC
pH-range	pH 6.5 – pH 9.5, highly reduced dependence on pH – value (see diagram last page “relative dependence on pH”)
Run-in time	First start-up approx. 2 h
Response time	T ₉₀ : approx. 2 min


	<h1>TARAline BR1</h1>	
Zero point adjustment	Not necessary	
Slope calibration	At the device, by analytical determination of the bromine concentration Recommendation depending on bromine agent: - Free bromine DPD1 - method - BCDMH DPD4 - method	
Cross sensitivities/ interferences	Cl ₂ : is also measured ClO ₂ : is also measured O ₃ : is also measured Corrosion inhibitors can lead to measuring errors. Stabilisers for water hardness can lead to measuring errors.	
Absence of the disinfectant	Max. 24 h	
Connection	analog-out/analog version: 4-pole plug adapter analog-out/digital version: 4-pole plug adapter digital-out/digital version: 5-pole M12, plug-on flange 4-20 mA version: 2-pole terminal or 5-pole M12, plug-on flange	
max. length of sensor cable (depending on internal signal processing)	analog	< 30 m
	digital	> 30 m are permissible Maximum cable length depends on application
material	Microporous hydrophilic membrane, PVC, PEEK ,stainless steel 1.4571	
Size	diameter: approx. 25 mm Length: BR1 (analog-out/analog) approx. 175 mm BR1 (analog-out/digital) approx. 195 mm BR1 (digital-out/digital) approx. 205 mm BR1 4-20 mA approx. 220 mm (2-pole-terminal) approx. 190 mm (5-pole-M12)	
Transport	+5 ... +50 °C (Sensor, electrolyte, membrane cap)	
storage	Sensor:	dry and without electrolyte no limit at +5 ... +40 °C
	Electrolyte:	in original bottle protected from sunlight at +5 ... +35 °C min. 1 year or until the specified EXP-Date
	Membrane cap:	in original packing no limit at +5 ... +40 °C (used membrane caps can not be stored)
maintenance	Regularly control of the measuring signal, min. once a week The following information highly depends on the water quality: Change of the membrane cap: once a year Change of the electrolyte: every 3 - 6 months	
	EMC-Testing DIN EN 61326-1, 61326-2-3, 63000 RoHS compliant	

Technical Data

1. BR1 (analog output, analog internal signal processing)

analog-out / analog

A potential-free electrical connection is necessary as the sensor electronic is not equipped with a galvanical isolation.


	Measuring range in ppm	Resolution in ppm	Output Output resistance	Nominal slope (at pH 7.2) in mV/ppm	Voltage supply	Connection
BR1H	0.005...2.000	0.001	analog 0...-2000 mV	-1000	±5 - ±15 VDC	4-pole screw connector
BR1N	0.05...20.00	0.01	1 kΩ	-100	10 mA	

(Subject to technical changes!)

2. BR1 (analog output, digital internal signal processing)

analog-out / digital

- The power supply is galvanically isolated inside of the sensor.
- The output signal is galvanically isolated too, that means potential-free.


	Measuring range in ppm	Resolution in ppm	Output Output resistance	Nominal Slope (at pH 7.2) in mV/ppm	Power supply	Connection
BR1H-An	0.005...2.000	0.001	analog 0...-2 V (max. -2.5 V)	-1000	9-30 VDC approx. 56-20 mA	4-pole screw connector
BR1N-An	0.05...20.00	0.01	1 kΩ	-100		
BR1H-Ap	0.005...2.000	0.001	analog 0...+2 V (max. +2.5 V)	+1000		
BR1N-Ap	0.05...20.00	0.01	1 kΩ	+100		

(Subject to technical changes!)

3. BR1 (digital output, digital internal signal processing)

digital-out / digital

- The power supply is galvanically isolated inside of the sensor.
- The output signal is galvanically isolated too, that means potential-free.

	Measuring range in ppm	Resolution in ppm	Output Output resistance	Power supply	Connection
BR1H-M0c	0.005...2.000	0.001	Modbus RTU	9-30 VDC	5-pole M12 plug-on flange
BR1N-M0c	0.05...20.00	0.01	There are no terminating resistors in the sensor.	approx. 56-20 mA	


(Subject to technical changes!)

4. BR1 4-20 mA (analog output, analog internal signal processing)

analog-out / analog


A potential-free electrical connection is necessary as the sensor electronic is not equipped with a galvanical isolation.

4.1 Electrical connection: 2 pole terminal clamp

	Measuring range in ppm	Resolution in ppm	Output Output resistance	Nominal slope (at pH 7.2) in mA/ppm	Voltage supply	Connection
BR1MA-2	0.005 ... 2.000	0.001	analog 4...20 mA uncalibrated	8.0	12...30 VDC R _L = 50Ω (12V)... 900Ω (30V)	2-pole terminal (2 x 1 mm ²) Recommended: Round cable ∅ 4 mm 2 x 0.34 mm ²
BR1MA-5	0.05 ... 5.00	0.01		3.2		
BR1MA-10	0.05 ... 10.00	0.01		1.6		
BR1MA-20	0.05 ... 20.00	0.01		0.8		

(Subject to technical changes!)

4.2 Electrical connection: 5 pole M12 plug-on flange

	Measuring range	Resolution	Output Output resistance	Nominal slope (at pH 7.2)	Voltage supply	Connection
	in ppm	in ppm		in mA/ppm		
BR1MA-2-M12	0.005 ... 2.000	0.001	analog 4...20 mA uncalibrated	8.0	12...30 VDC R _L = 50Ω (12V)... 900Ω (30V)	5-pole M12 plug- on flange Function of wires: PIN2: +U PIN3: -U
BR1MA-5-M12	0.05 ... 5.00	0.01		3.2		
BR1MA-10-M12	0.05 ... 10.00	0.01		1.6		
BR1MA-20-M12	0.05 ... 20.00	0.01		0.8		

(Subject to technical changes!)

Spare Parts

Type	Membrane cap	Electrolyte	Emery	O-ring
All BR1	M48.2 Art. No. 11047	ECP1.4/GEL, 100 ml Art. No. 11006.1	S1 Art. No. 11908	14 x 1.8 NBR Art. No. 11806

(Subject to technical changes!)

relative dependence on pH

Temperature: 25°C / Flow rate: 30 l/h

